

Scope of Accreditation For Instrulab, Inc.

1205 Lamar Street
Dayton, OH 45404-0098
Stanley I. Tekamp
937-223-2241

In recognition of a successful assessment to ISO/IEC 17025:2005, accreditation is granted to **Instrulab, Inc.** to perform the following Calibrations:

Accreditation granted through: **November 11, 2013**

Calibration

Electricity and Magnetism – Resistance

Calibration Parameter/Equipment ¹	Range	Calibration and Measurement Capability(+/-) ²	Remarks
Resistance, Source ³	1 Ω 10 Ω 25 Ω 50 Ω 75 Ω 100 Ω	0.0001 Ω 0.0001 Ω 0.0001 Ω 0.0002 Ω 0.0003 Ω 0.0004 Ω	Six Standard Resistors
	(0 to 300) Ω	0.0008 % of reading	
Resistance, Measure ³	(0 to 16 000) Ω	0.0004 % of reading	L & N 8079 Bridge
100Ω PRT 385 Simulate ⁴	(-150 to 750) °F	0.4 °F	FGH Ezeal 5 Process Calibrator Reference IEC 751
Thermistor, Simulate (YSI 400 Series 2252 Ω @ 25°C) ³	(0 to 80) °C	0.002 °C	Decade Resistance Box
	(80 to 100) °C	0.003 °C	

Time and Frequency – Time Dissemination

Calibration Parameter/Equipment ^{1,3}	Range	Calibration and Measurement Capability(+/-) ²	Remarks
Timer	(0 to 24) hr	2.0 sec / 24 hr	Robic Chronometer

Mass – Pressure

Calibration Parameter/Equipment ^{1,3}	Range	Calibration and Measurement Capability(+/-) ²	Remarks
Low Pressure	(0 to 72) psi	0.20 % of reading + 1 LSD	Meriam Pressure Calibrator
Vacuum	(0 to 30) inHg	0.20 % of reading + 1 LSD	

Thermodynamics – Thermocouples

Calibration Parameter/Equipment ¹	Range	Calibration and Measurement Capability(+/-) ²	Remarks
Thermocouple Measure ³ Type J	(-100 to 500) °F	0.2 °F	Comparison technique using 100Ω PRT in liquid bath or dry block calibrators and type “S” thermocouple in calibration furnace.
	(500 to 701) °F	0.3 °F	
	(702 to 1 300) °F	1.4 °F	
	(1 300 to 1 500) °F	1.7 °F	
	Type K	(-100 to 100) °F	
(100 to 500) °F		0.2 °F	
(500 to 701) °F		0.3 °F	
(702 to 1 300) °F		1.4 °F	
(1 300 to 1 700) °F		1.7 °F	
Type N	(-320 to 300) °F	0.2 °F	
	(300 to 701) °F	0.3 °F	
	(702 to 1 100) °F	1.4 °F	
	(1 100 to 2 000) °F	1.8 °F	
Type T	(-320 to 100) °F	0.3 °F	
	(100 to 700) °F	0.2 °F	
	700 °F	0.3 °F	
Type S/R	(100 to 500) °F	0.2 °F	
	(500 to 701) °F	0.3 °F	
	(702 to 1300) °F	1.4 °F	
	(1 300 to 2 000) °F	1.8 °F	
100Ω Platinum RTD Measure ³	(-200 to -100) °C	0.028 °C	Distilled water ice bath and comparison technique using a 100Ω PRT in liquid baths and dry block calibrators.
	(-100 to -70) °C	0.036 °C	
	(-70 to 0) °C	0.029 °C	
	0 °C	0.009 °C	
	(0 to 150) °C	0.017 °C	
	(150 to 200) °C	0.020 °C	
	(200 to 250) °C	0.023 °C	
	(250 to 300) °C	0.029 °C	
	(300 to 350) °C	0.032 °C	
	(350 to 400) °C	0.035 °C	
400 °C	0.043 °C		
Temperature Source ³	-196 °C	0.022 °C	Liquid Nitrogen, Methanol, Dry Ice, Distilled Water Ice Bath and 100Ω PRT in Dry Block Calibrator
	(-70 to 0) °C	0.012 °C	
	0 °C	0.009 °C	
	(0 to 100) °C	0.013 °C	
	(100 to 150) °C	0.014 °C	
	(150 to 200) °C	0.016 °C	
	(200 to 250) °C	0.018 °C	
	(250 to 300) °C	0.025 °C	
	(300 to 350) °C	0.026 °C	
	(350 to 400) °C	0.029 °C	
400 °C	0.037 °C		
Thermistor, Measure (YSI 400 Series 2252 Ohms @ 25 °C) ³	0 °C	0.012 °C	Distilled Water Ice Bath and 100Ω PRT in Liquid Bath
	(1 to 40) °C	0.025 °C	
	(40 to 60) °C	0.027 °C	
	(60 to 80) °C	0.031 °C	
	(80 to 100) °C	0.034 °C	

Electricity and Magnetism – Electrical Temperature Simulation

Calibration Parameter/Equipment ¹	Range	Calibration and Measurement Capability(+/-) ²	Remarks	
Thermocouple Simulation ³				
Type J	(-150 to 2 000) °F	0.2 °F	Precision Potentiometer with Distilled Water Ice Bath and Appropriate Thermocouple Wire	
Type K	(-150 to 1 500) °F (1 500 to 2 500) °F	0.3 °F 0.4 °F		
Type T	(-320 to 750) °F	0.2 °F		
Type R	(500 to 2 500) °F (2 500 to 3 200) °F	0.6 °F 0.7 °F		
Type S	(500 to 2 500) °F (2 500 to 3 200) °F	0.6 °F 0.7 °F		
Type B	(500 to 750) °F (750 to 3 300) °F	0.5 °F 0.4 °F		
Type N	(-320 to 1 500) °F (1 500 to 2 370) °F	0.3 °F 0.4 °F		Reference ASTM E230
Type E	(-150 to 1 600) °F (1 600 to 1 830) °F	0.2 °F 0.3 °F		
Type C 5/26	(500 to 1 000) °F (1 000 to 2 500) °F (2 500 to 3 500) °F (3 500 to 4 000) °F 4 000 °F	2.8 °F 2.9 °F 3.0 °F 3.2 °F 3.4 °F		
Thermocouple Simulate ⁴ or Measure	(-150 to 0) °F (0 to 1 000) °F	1.1 °F 0.9 °F		FGH Ezecal 5 Process Calibrator
Type E	(1 000 to 1 500) °F (1 500 to 1 830) °F 1 830 °F	1.1 °F 1.3 °F 1.5 °F		
Type J	(-150 to 0) °F (0 to 250) °F (250 to 1 000) °F (1 000 to 1 500) °F (1 500 to 2 000) °F 2 000 °F	1.1 °F 0.6 °F 0.9 °F 1.1 °F 1.3 °F 1.5 °F		
Type K	(-100 to 0) °F (0 to 750) °F (750 to 1 500) °F (1 500 to 2 000) °F (2 000 to 2 500) °F 2 500 °F	1.1 °F 0.9 °F 1.1 °F 1.3 °F 1.5 °F 1.7 °F		
Type T	(-320 to -150) °F (-150 to 0) °F (0 to 750) °F 750 °F	1.3 °F 1.1 °F 0.9 °F 1.1 °F		
			Reference ASTM E230	

Calibration Parameter/Equipment ¹	Range	Calibration and Measurement Capability(+/-) ²	Remarks
Type N	(-320 to -150) °F	1.3 °F	FGH Ezeal 5 Process Calibrator Reference ASTM E230
	(-150 to 0) °F	1.1 °F	
	(0 to 1 000) °F	0.9 °F	
	(1 000 to 1 500) °F	1.1 °F	
	(1 500 to 2 000) °F	1.3 °F	
	(2 000 to 2 370) °F	1.5 °F	
	2 370 °F	1.7 °F	
Type R	(0 to 1 000) °F	2.8 °F	
	(1 000 to 1 500) °F	3.0 °F	
	(1 500 to 2 000) °F	3.2 °F	
	(2 000 to 2 500) °F	3.4 °F	
	(2 500 to 3 000) °F	3.6 °F	
	(3 000 to 3 200) °F	3.7 °F	
	3 200 °F	3.9 °F	
Type S	(0 to 1 000) °F	2.8 °F	
	(1 000 to 1 500) °F	3.0 °F	
	(1 500 to 2 000) °F	3.2 °F	
	(2 000 to 2 500) °F	3.4 °F	
	(2 500 to 3 000) °F	3.6 °F	
	(3 000 to 3 200) °F	3.8 °F	
	3 200 °F	3.9 °F	
Type B	(500 to 1 000) °F	2.9 °F	
	(1 000 to 1 500) °F	3.1 °F	
	(1 500 to 2 000) °F	3.3 °F	
	(2 000 to 2 500) °F	3.5 °F	
	(2 500 to 3 000) °F	3.7 °F	
	(3 000 to 3 300) °F	3.9 °F	
	3 300 °F	4.1 °F	
Type C	(500 to 1 000) °F	6.7 °F	
	(1 000 to 1 500) °F	7.1 °F	
	(1 500 to 2 000) °F	7.4 °F	
	(2 000 to 2 500) °F	7.8 °F	
	(2 500 to 3 000) °F	8.2 °F	
	(3 000 to 3 500) °F	8.5 °F	
	(3 500 to 4 000) °F	8.9 °F	
4 000 °F	9.3 °F		

Notes:

- 1) Laboratory offers calibration services at the laboratory's own facilities and at the client or other agreed upon facilities.
- 2) Calibration and Measurement Capability represents expanded uncertainties at approximately a 95% confidence level using a coverage factor of k=2.
- 3) Parameters that are to be calibrated in-house.
- 4) Parameters that are to be calibrated on-site.

 Approved by: 

 R. Douglas Leonard Jr.
Chief Technical Officer

 Date: November 11, 2010

Re-Issued: 11/11/10